

AK

**Notice of Allowability**

Application No.

10/634,315

Examiner

Allen C. Ho

Applicant(s)

HIXSON, GORDON L.

Art Unit

2882

**-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--**

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to amendment filed on 16 May 2005.
2. ☒ The allowed claim(s) is/are 1-22.
3. ☒ The drawings filed on 05 August 2003 are accepted by the Examiner.
4. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
  - a) ☐ All    b) ☐ Some\*    c) ☐ None    of the:
    1. ☐ Certified copies of the priority documents have been received.
    2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
    3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

\* Certified copies not received: \_\_\_\_\_.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.  
**THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.**

5. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
6. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
  - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
    - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date \_\_\_\_\_.
  - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date \_\_\_\_\_.

Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
7. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

**Attachment(s)**

- |  |   |
|--|---|
| 1. <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)  | 5. <input type="checkbox"/> Notice of Informal Patent Application (PTO-152)                       |
| 2. <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                                 | 6. <input checked="" type="checkbox"/> Interview Summary (PTO-413),<br>Paper No./Mail Date _____. |
| 3. <input type="checkbox"/> Information Disclosure Statements (PTO-1449 or PTO/SB/08),<br>Paper No./Mail Date _____. | 7. <input checked="" type="checkbox"/> Examiner's Amendment/Comment                               |
| 4. <input type="checkbox"/> Examiner's Comment Regarding Requirement for Deposit<br>of Biological Material           | 8. <input checked="" type="checkbox"/> Examiner's Statement of Reasons for Allowance              |
|  | 9. <input type="checkbox"/> Other _____.  |

sk

### EXAMINER'S AMENDMENT

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with Stephen J. Stark (Reg. No. 43,152) on 09 June 2005.

Claims 1 and 20 have been amended as follows:

1. (Currently Amended) A mammographic compression device comprising:

a substantially planar image detector extending along a horizontal plane;

a compression paddle selectively moveable vertically toward and away from the image detector along a guide, said compression paddle having

a first planar compression surface extending from ~~the~~a chest wall end of the compression paddle, said first planar compression surface extending less than about one inch from the chest wall end;

a third compression surface spaced no more than an inch from the first planar compression surface, said third planar compression surface forming an angle relative to the first compression surface in a range of one hundred seventy three to one hundred seventy nine degrees; said third planar compression surface extending downwardly angled~~relative to the first planar compression surface~~ toward the image detector.

20. (Currently Amended) A mammographic compression device comprising:

Art Unit: 2882

a breast support plate; and

a compression paddle having a lip at a chest wall end of the paddle, a first compression surface beginning at the chest wall end of the paddle, said first compression surface being concave relative to the breast support plate and extending less than about 2 inches from the chest wall end of the paddle, and a second planar compression surface extending from the first concave compression surface to a nipple end of the compression paddle wherein a plane extending through the second planar compression surface forms an angle with the lip of between about 92 and about 98 degrees so that the second ~~phase~~planar compression surface is downwardly oriented toward the breast support plate relative to the first compression surface.

***Allowable Subject Matter***

2. Claims 1-22 are allowed.
3. The following is an examiner's statement of reasons for allowance:

The prior art fails to teach or fairly suggest a compression paddle that comprises a plurality of compression surfaces, the compression surfaces being angled relative to each other, and/or being concave as claimed.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Art Unit: 2882

*Conclusion*

4. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure:

Barbarisi (U. S. Patent No. 4,943,986) disclosed a mammography compression paddle.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Allen C. Ho whose telephone number is (571) 272-2491. The examiner can normally be reached on Monday - Friday from 8:00 am - 5:00 pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Edward J. Glick can be reached at (571) 272-2490. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



Allen C. Ho  
Primary Examiner  
Art Unit 2882

09 June 2005